



Certificate of Traceability

Level I Laboratory Certification

This certifies that **Labor Dr. Wisplinghoff / Laborgemeinschaft Köln**, using **Tosoh G11** has participated in and successfully completed NGSP Level I Laboratory certification and is traceable to the **Diabetes Control and Complications Trial** Reference method. The comparison was performed with: **European Reference Laboratory ESRL#10**

The system evaluated was:

Instrument: G11	Calibrator Lot: ZB 3010, ZB3010	Column Lot: KB1154N
Reagent Lot: N1-101K, N1-207H, 1-301K, HW085D/B	Calibrator Assigned Values: 40.60, 96.40 mmol/mol	

Date of Certification: November 1, 2023

Certification Expires: November 1, 2024

Handwritten signature of David Sachs in black ink.

NGSP Steering Committee Chair

Handwritten signature of Randie R Little PhD in black ink.

NGSP Network Coordinator

Handwritten signature of the SRL director/supervisor in black ink.

SRL director/ supervisor